

SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/26/2020

OPERATOR NAME:  
HUNTER A NERBER  
Permit #: 48077  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:40
INTERNAL STANDARD	VERIFIED	06:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:40
BLANK TEST	0.00	06:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:42
BLANK TEST	0.00	06:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:43
BLANK TEST	0.00	06:44



SUPERVISOR TEST



SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/20/2020

OPERATOR NAME:  
HUNTER A NERBER  
Permit #: 48077  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:48
INTERNAL STANDARD	VERIFIED	06:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:48
BLANK TEST	0.00	06:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:49
BLANK TEST	0.00	06:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:51
BLANK TEST	0.00	06:52



SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/14/2020

OPERATOR NAME:  
HUNTER A NERBER  
Permit #: 48077  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:57
INTERNAL STANDARD	VERIFIED	10:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:00
BLANK TEST	0.00	11:01



SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/10/2020

OPERATOR NAME:  
HUNTER A NERBER  
Permit #: 48077  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:58
INTERNAL STANDARD	VERIFIED	06:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:00
BLANK TEST	0.00	07:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02



SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/04/2020

OPERATOR NAME:  
DEREK J CZEROW  
Permit #: 43605  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:54
INTERNAL STANDARD	VERIFIED	13:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58



SOLUTION CHANGE

SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 02/01/2020

OPERATOR NAME:  
DEREK J CZEROW  
Permit #: 43605  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:02

**ALCOHOL REFERENCE SOLUTION  
FOR SIMULATOR**

19250  
LOT NO.




3/30/2020  
EXP. DATE

1131  
BOT. NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH  
A PROPERLY OPERATING BREATH SIMULATOR AND  
BREATH ALCOHOL ANALYZER, THE SOLUTION WILL  
PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.  
HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06 *D.J.C.* 1/5/2020 #6034



SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 01/27/2020

OPERATOR NAME:  
HUNTER A NERBER  
Permit #: 48077  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:24
INTERNAL STANDARD	VERIFIED	06:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	06:27
BLANK TEST	0.00	06:28



SUPERVISOR TEST



SOLVAY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 138506

Date: 01/20/2020

OPERATOR NAME:  
 HUNTER A NERBER  
 Permit #: 48077  
 Operator Agency/Dept: SOLVAY VILLAGE PD  
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:47
INTERNAL STANDARD	VERIFIED	22:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:48
BLANK TEST	0.00	22:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:50
BLANK TEST	0.00	22:50





SUPERVISOR TEST




SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 01/18/2020

OPERATOR NAME:  
DEREK J CZEROW  
Permit #: 43605  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:07
INTERNAL STANDARD	VERIFIED	07:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11



SUPERVISOR TEST




SOLVAY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 138506

Date: 01/12/2020

OPERATOR NAME:  
 DEREK J CZEROW  
 Permit #: 43605  
 Operator Agency/Dept: SOLVAY VILLAGE PD  
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:26
INTERNAL STANDARD	VERIFIED	07:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:28
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:29
BLANK TEST	0.00	07:30



SOLUTION CHANGE

SOLVAY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 138506

Date: 01/05/2020

OPERATOR NAME:  
DEREK J CZEROW  
Permit #: 43605  
Operator Agency/Dept: SOLVAY VILLAGE PD  
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:59
INTERNAL STANDARD	VERIFIED	06:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	07:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03

ALCOHOL REFERENCE SOLUTION  
FOR SIMULATOR

19250  
LOT NO.

3/30/2020  
EXP. DATE

44223  
BOT NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH  
A PROPERLY OPERATING BREATH SIMULATOR AND  
BREATH ALCOHOL ANALYZER, THE SOLUTION WILL  
PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.  
HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

*DJS* 2012 12/27

