

2019

**DATAMASTER
RECORD OF INSPECTION
MAINTENANCE
CALIBRATION**



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 138506, performed on March 28, 2019, by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 4/1/2019 12:13 PM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

BREATH TEST INSTRUMENT RECORD OF INSPECTION / MAINTENANCE / CALIBRATION

Name of Submitting Agency: Solvay Village Police Department
 Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT
 Date: March 28, 2019 Serial Number: 138506

Reference Solution Lot Number: 18371
 Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

| Test No | Test Result | Simulator Temperature |
|---------|-------------|--------------------------|
| 1 | 0.098 % | 34.01 degrees centigrade |
| 2 | 0.098 % | 34.01 degrees centigrade |
| 3 | 0.098 % | 34.01 degrees centigrade |
| 4 | 0.098 % | 34.01 degrees centigrade |

Reference Solution Lot Number: 18161
 Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

| Test No | Test Result | Simulator Temperature |
|---------|-------------|--------------------------|
| 1 | 0.079 % | 34.01 degrees centigrade |
| 2 | 0.080 % | 34.01 degrees centigrade |
| 3 | 0.080 % | 34.01 degrees centigrade |
| 4 | 0.080 % | 34.01 degrees centigrade |


Reference Solution Lot Number: 18280
 Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

| Test No | Test Result | Simulator Temperature |
|---------|-------------|--------------------------|
| 1 | 0.182 % | 34.01 degrees centigrade |
| 2 | 0.183 % | 34.01 degrees centigrade |
| 3 | 0.182 % | 34.01 degrees centigrade |
| 4 | 0.183 % | 34.01 degrees centigrade |

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 138506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
 ESRA by Martin Schilling
 on 3/28/2019 2:19 PM

Martin Schilling
 Highway Safety Equipment Technician
 Breath Analysis Operator Permit # 45926

Reviewed by:




Digitally signed under
 ESRA by Frank Sgarlata
 on 4/1/2019 9:03 AM

Frank Sgarlata
 Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 138506
Date: March 28, 2019

- 1. General Appearance and condition of this device -----: OK
- 2. Check controls and indicators -----: OK
- 3. Check breath flow system -----: OK
- 4. Check printer operation -----: OK
- 5. System cycles through all modes -----: OK



6. Record any repairs made to this instrument and list services provided: **INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION**

Performed by:

Digitally signed under
ESRA by Martin Schilling
on 3/28/2019 2:19 PM

**Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926**

Reviewed by:

Digitally signed under
ESRA by Frank Sgarlata
on 4/1/2019 9:03 AM

**Frank Sgarlata
Breath Analysis Technical Supervisor**



**Division of Criminal
Justice Services**

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 138506, performed on November 06, 2019 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance/calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under ESRA
by Joshua Vinehout
on 11/13/2019 11:02 AM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT
Serial Number: 138506
Date: November 06, 2019


Reference Solution Lot Number: 19180
Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

| Test No | Test Result | Simulator Temperature |
|---------|-------------|--------------------------|
| 1 | 0.100 % | 33.99 degrees centigrade |
| 2 | 0.100 % | 33.99 degrees centigrade |
| 3 | 0.101 % | 33.99 degrees centigrade |
| 4 | 0.101 % | 33.99 degrees centigrade |

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 138506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.


Performed by:



Digitally signed under
ESRA by Joseph Iavarone
on 11/6/2019 3:45 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:




Digitally signed under
ESRA by Frank Sgarlata
on 11/12/2019 10:50 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 138506
Date: November 06, 2019

1. General Appearance and condition of this device -----: N/A
2. Check controls and indicators -----: OK
3. Check breath flow system -----: OK
4. Check printer operation -----: N/A
5. System cycles through all modes -----: OK
6. Record any repairs made to this instrument and list services provided: **CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.**



Performed by:

Joseph Iavarone

Digitally signed under
ESRA by Joseph Iavarone
on 11/6/2019 3:45 PM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 44106

Reviewed by:

Frank Sgarlata

Digitally signed under
ESRA by Frank Sgarlata
on 11/12/2019 10:50 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

SOLVAY VILLAGE
POLICE DEPT
DATAMASTER dmt: 138506

Date: 11/06/2019

Nominal = 0.100 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.000 | 09:17 |
| INTERNAL STANDARD | VERIFIED | 09:17 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.100 | 09:18 |
| BLANK TEST | 0.000 | 09:19 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.100 | 09:19 |
| BLANK TEST | 0.000 | 09:20 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.101 | 09:20 |
| BLANK TEST | 0.000 | 09:21 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.101 | 09:22 |
| BLANK TEST | 0.000 | 09:22 |

Average = 0.1005

Std Dev = 0.0006

Supervisor Record



Control Test Record

Serial Number: 138506

Test Date: 11/06/2019 07:40:19

Location: 190

SOLVAY VILLAGE POLICE DEPT.

Supervisor Information

First Name: Derek J

Last Name: Czerow

Permit: 43605

Agency: SOLVAY VILLAGE PD

Simulator Information

Lot: 19180 Nominal: 0.100

Test Results


Alcohol 1: 0.10 33.99

Alcohol 2: 0.10 33.99

Alcohol 3: 0.10 33.99

Solution Change:

Plot 1 of 3 <=>



Status

Status Code:

Date Time:

Message: