

2022

DATAMASTER

RECORD OF INSPECTION

MAINTENANCE

CALIBRATION

NOVEMBER 2022

11/07/2022



**Division of Criminal
Justice Services**

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NATL PATENT ANALYTICAL SYSTEM, DATAMASTER DMT breath test instrument, serial number 138506, performed on November 07, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 11/18/2022 10:34 AM

**Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety**

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 138506
Date: November 07, 2022

Reference Solution Lot Number : 22120
Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.00 degrees centigrade
2	0.099 %	34.00 degrees centigrade
3	0.099 %	34.01 degrees centigrade
4	0.099 %	34.00 degrees centigrade

Reference Solution Lot Number : 22210
Reference Solution Standard : 0.08% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.081 %	34.00 degrees centigrade
2	0.080 %	34.01 degrees centigrade
3	0.081 %	34.00 degrees centigrade
4	0.080 %	34.00 degrees centigrade


Reference Solution Lot Number : 22060
Reference Solution Standard : 0.18% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.180 %	34.00 degrees centigrade
2	0.180 %	34.00 degrees centigrade
3	0.181 %	34.01 degrees centigrade
4	0.181 %	34.00 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NATL PATENT ANALYTICAL SYSTEM breath test instrument model DATAMASTER DMT, serial number 138506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.


I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under ESRA
by George Coffin
on 11/07/2022 03:17 PM

Reviewed by:



Digitally signed under ESRA
by Frank Sgarlata
on 11/07/2022 03:20 PM

George Coffin
Highway Safety Equipment Technician
Breath Analysis Operator Permit #48534

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 138506
Date: November 07, 2022

- 1. General appearance and condition of this device -----: OK

- 2. Check controls and indicators -----: OK

- 3. Check breath flow system -----: OK

- 4. Check printer operation -----: OK

- 5. System cycles through all modes -----: OK



6. Record any repairs made to this instrument and list services provided : INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.

Performed by:

Digitally signed under ESRA
by George Coffin
on 11/07/2022 03:17 PM

Reviewed by:

Digitally signed under ESRA
by Frank Sgarlata
on 11/07/2022 03:20 PM

George Coffin
Highway Safety Equipment Technician
Breath Analysis Operator Permit #48534

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

SOLVAY VILLAGE
POLICE DEPT
DATAMASTER dmt: 138506





Date: 11/07/2022

OPERATOR NAME:
GEORGE R COFFIN
Permit #: 48534
Operator Agency/Dept: SOLVAY VILLAGE PD
Reference Standard #: 22210

Nominal = 0.080 g/210L

BLANK TEST	0.000	12:49
INTERNAL STANDARD	VERIFIED	12:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.081	12:49
BLANK TEST	0.000	12:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.080	12:50
BLANK TEST	0.000	12:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.081	12:52
BLANK TEST	0.000	12:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.080	12:53
BLANK TEST	0.000	12:54

Average = 0.0805
Std Dev = 0.0006



TECHNICIAN TEST

SOLVAY VILLAGE
POLICE DEPT
DATAMASTER dmt: 138506


Date: 11/07/2022

OPERATOR NAME:
GEORGE R COFFIN
Permit #: 48534
Operator Agency/Dept: SOLVAY VILLAGE PD
Reference Standard #: 22120

Nominal = 0.100 g/210L

BLANK TEST	0.000	12:56
INTERNAL STANDARD	VERIFIED	12:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	12:57
BLANK TEST	0.000	12:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	12:58
BLANK TEST	0.000	12:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.099	12:59
BLANK TEST	0.000	13:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.099	13:01
BLANK TEST	0.000	13:01

Average = 0.0990
Std Dev = 0.0000



TECHNICIAN TEST

SOLVAY VILLAGE
POLICE DEPT
DATAMASTER dmt: 138506





Date: 11/07/2022

OPERATOR NAME:
GEORGE R COFFIN
Permit #: 48534
Operator Agency/Dept: SOLVAY VILLAGE PD
Reference Standard #: 22060

Nominal = 0.180 g/210L

BLANK TEST	0.000	13:04
INTERNAL STANDARD	VERIFIED	13:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.180	13:05
BLANK TEST	0.000	13:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.180	13:06
BLANK TEST	0.000	13:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.181	13:07
BLANK TEST	0.000	13:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.181	13:09
BLANK TEST	0.000	13:10

Average = 0.1805
Std Dev = 0.0006



2022
DATAMASTER
RECORD OF INSPECTION
MAINTENANCE
CALIBRATION

MAY 2022

Brought in for repair

CALIBRATED

05/23/2022



**Division of Criminal
Justice Services**

**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, Steven J. Carluccio, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NATL PATENT ANALYTICAL SYSTEM, DATAMASTER DMT breath test instrument, serial number 138506, performed on May 12, 2022 by Office of Public Safety employee Joseph Iavarone, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 05/23/2022 11:53 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 138506
Date: May 12, 2022


Reference Solution Lot Number : 22030
Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.101 %	33.99 degrees centigrade
2	0.101 %	33.99 degrees centigrade
3	0.101 %	33.99 degrees centigrade
4	0.101 %	33.99 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NATL PATENT ANALYTICAL SYSTEM breath test instrument model DATAMASTER DMT, serial number 138506, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.


I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under ESRA
by Joseph Iavarone
on 05/12/2022 02:47 PM

Reviewed by:



Digitally signed under ESRA
by Frank Sgarlata
on 05/13/2022 06:32 AM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit #44106

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Solvay Village Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 138506
Date: May 12, 2022

- 1. General appearance and condition of this device -----: N/A

- 2. Check controls and indicators -----: OK

- 3. Check breath flow system -----: OK

- 4. Check printer operation -----: N/A

- 5. System cycles through all modes -----: OK

- 6. Record any repairs made to this instrument and list services provided : CHECKED REMOTE OPERATION, RAN DIAGNOSTIC AND TECHNICIAN TESTS, VERIFIED CALIBRATION.



Performed by:

Joseph Iavarone

Digitally signed under ESRA
by Joseph Iavarone
on 05/12/2022 02:47 PM

Reviewed by:

Frank Sgarlata

Digitally signed under ESRA
by Frank Sgarlata
on 05/13/2022 06:32 AM

Joseph Iavarone
Highway Safety Equipment Technician
Breath Analysis Operator Permit #44106

Frank Sgarlata
Breath Analysis Technical Supervisor

SOLVAY VILLAGE
POLICE DEPT
DATAMASTER dmt: 136506

Date: 05/12/2022

Nominal = 0.100 g/210L

BLANK TEST	0.000	14:28
INTERNAL STANDARD	VERIFIED	14:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.101	14:29
BLANK TEST	0.000	14:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.101	14:30
BLANK TEST	0.000	14:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.101	14:31
BLANK TEST	0.000	14:32
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.101	14:32
BLANK TEST	0.000	14:33

Average = 0.1010
Std Dev = 0.0000

Control Test Record

Serial Number: 138506
Test Date: 05/09/2022 16:04:27
Location: 190
SOLVAY VILLAGE POLICE DEPT.

Simulator Information

Lot: 22030 Nominal: 0.100

Test Results


Alcohol 1: 0.10 33.99
Alcohol 2: 0.10 33.99
Alcohol 3: 0.10 33.99

Supervisor Information

First Name: H A
Last Name: Nerber
Permit: 48077
Agency: SOLVAY VILLAGE PD

Solution Change:

Plot 1 of 3



Status

Status Code:
Date Time:
Message: